

Journal of Biomedical Optics

SPIEDigitalLibrary.org/jbo

Test-retest assessment of cortical activation induced by repetitive transcranial magnetic stimulation with brain atlas-guided optical topography

Fenghua Tian
F. Andrew Kozel
Amarnath Yennu
Paul E. Croarkin
Shawn M. McClintock
Kimberly S. Mapes
Mustafa M. Husain
Hanli Liu

Test-retest assessment of cortical activation induced by repetitive transcranial magnetic stimulation with brain atlas-guided optical topography

Fenghua Tian,^a F. Andrew Kozel,^{b,c} Amarnath Yennu,^a Paul E. Croarkin,^{c,d} Shawn M. McClintock,^c Kimberly S. Mapes,^c Mustafa M. Husain,^c and Hanli Liu^a

^aUniversity of Texas at Arlington, Department of Bioengineering, 500 UTA Building, Arlington, Texas

^bUniversity of South Florida, Department of Psychiatry and Behavioral Neurosciences, Tampa, Florida

^cUniversity of Texas Southwestern Medical Center, Department of Psychiatry, Dallas, Texas

^dMayo Clinic, Department of Psychiatry and Psychology, Rochester, Minnesota

Abstract. Repetitive transcranial magnetic stimulation (rTMS) is a technology that stimulates neurons with rapidly changing magnetic pulses with demonstrated therapeutic applications for various neuropsychiatric disorders. Functional near-infrared spectroscopy (fNIRS) is a suitable tool to assess rTMS-evoked brain responses without interference from the magnetic or electric fields generated by the TMS coil. We have previously reported a channel-wise study of combined rTMS/fNIRS on the motor and prefrontal cortices, showing a robust decrease of oxygenated hemoglobin concentration ($\Delta[\text{HbO}_2]$) at the sites of 1-Hz rTMS and the contralateral brain regions. However, the reliability of this putative clinical tool is unknown. In this study, we develop a rapid optical topography approach to spatially characterize the rTMS-evoked hemodynamic responses on a standard brain atlas. A hemispherical approximation of the brain is employed to convert the three-dimensional topography on the complex brain surface to a two-dimensional topography in the spherical coordinate system. The test-retest reliability of the combined rTMS/fNIRS is assessed using repeated measurements performed two to three days apart. The results demonstrate that the $\Delta[\text{HbO}_2]$ amplitudes have moderate-to-high reliability at the group level; and the spatial patterns of the topographic images have high reproducibility in size and a moderate degree of overlap at the individual level. © 2012 Society of Photo-Optical Instrumentation Engineers (SPIE). [DOI: 10.1117/1.JBO.17.11.116020]

Keywords: repetitive transcranial magnetic stimulation; functional near-infrared spectroscopy; topography; oxygenated hemoglobin; deoxygenated hemoglobin; reliability; individual analysis.

Paper 12517 received Aug. 12, 2012; revised manuscript received Oct. 2, 2012; accepted for publication Oct. 9, 2012; published online Nov. 7, 2012.

1 Introduction

Transcranial magnetic stimulation (TMS) is a technology that has been traditionally used to study brain function as well as being a therapeutic tool. TMS produces rapidly changing magnetic pulses from an electromagnetic coil placed on the scalp. The magnetic pulses pass relatively unimpeded to the cortex and generate electrical fields that can depolarize neurons.¹ In recent years, repetitive TMS (rTMS) has been tested as a therapeutic tool for various neuropsychiatric disorders.^{1,2} For example, daily rTMS to the prefrontal cortex has been found to have antidepressant properties in patients with depression.^{3–9} In these clinical applications, objective measures of the impact of rTMS on the brain are essential to evaluate and optimize treatment. Numerous attempts to measure brain activity during TMS have been made with various neuroimaging modalities. Although functional MRI, EEG, and MEG have the temporal resolution to measure rapid brain changes with rTMS, each of these modalities measures brain activity by detecting either electrical or magnetic signals from the brain. The use of these neuroimaging techniques in rTMS treatment is limited by the coproduction of significant measurement artifacts due to electromagnetic

coupling or interference between the rTMS magnetic field and detection electromagnetic field.

Functional near-infrared spectroscopy (fNIRS) is an optical imaging technology that measures the light attenuation of the brain in a near infrared spectrum (670 to 900 nm).^{10,11} The near-infrared light is mainly absorbed by the oxygenated hemoglobin (HbO_2) and the deoxygenated hemoglobin (Hb) in cerebral blood flow. By measuring the change of light attenuation from a baseline state at two or more wavelengths, the changes of HbO_2 and Hb concentrations ($\Delta[\text{HbO}_2]$ and $\Delta[\text{Hb}]$) can be quantified.¹² Hemodynamic changes in the brain have been demonstrated to be tightly coupled with neuronal activations.¹³ Because fNIRS measures optical signals with magnetically compatible fibers, it does not interact with the magnetic or electric fields produced by the TMS coil. Thus fNIRS is a suitable tool to study the mechanisms of TMS, and a few reports have been published.^{14–19} Recently, we studied the motor cortex and the prefrontal cortex in response to rTMS at a frequency of 1 Hz and an intensity of 120% of the resting motor threshold (RMT).²⁰ For the first time, hemodynamic changes both under the TMS coil and contralateral to the coil were simultaneously measured with fNIRS. A robust decrease in $\Delta[\text{HbO}_2]$ was observed at the sites of rTMS and the contralateral brain regions. The temporal similarity between the ipsilateral and contralateral

Address all correspondence to: Hanli Liu, University of Texas at Arlington, Department of Bioengineering, 500 UTA Building, Arlington, Texas 76019-19138. E-mail: hanli@uta.edu

responses at each cortical location suggested strong interhemispheric connectivity during rTMS.

To date, these pilot studies have demonstrated the great potential of fNIRS to objectively assess the reactivity and connectivity of the cortex associated with rTMS. In order to have confidence in the combined rTMS/fNIRS technology as a useful clinical tool, reliability of the measurement must be established at the individual level (i.e., being reliable to assess a single patient's brain circuitry).²¹ However, there exists little rigorous evaluation of the test-retest reliability of the combined rTMS/fNIRS. In the literature, a few studies on the test-retested reliability of fNIRS as a stand-alone technology have been reported: an early study conducted by Watanabe et al.²² explored the repeated fNIRS measures in five subjects who underwent cognitive tasks and hyperventilation. Channel-wise hemodynamic changes between two sessions were assessed by intraclass correlation coefficients (ICCs) that demonstrated the acceptable reliability of fNIRS. The sample size in this study was relatively small, and the test-retest interval was largely different among subjects. Later studies conducted by Plitcha et al. included 12 subjects who underwent visual²³ and motor²⁴ stimulations with a constant test-retest interval of three weeks. Both the reliability of hemodynamic changes at the group level and the individual reproducibility of activated channels in terms of the size and the overlap were assessed. It demonstrated overall good to excellent reproducibility at the group level, whereas the results at the individual level were less robust. Similar methods were also used by Zhang et al. on NIRS-based resting-state functional connectivity.²⁵ All of these studies assessed the reliability of stand-alone fNIRS based on single channel or a cluster of channels in the absence of the anatomical base of the brain. Even though the locations of the channels on the brain could be identified with additional MRI, such as performed by Zhang et al., the detailed information were lacking with regard to the cortical regions in-between the channels.

In the present paper, first we have developed a rapid topography approach to characterize the rTMS-evoked hemodynamic responses on a standard brain atlas. The fNIRS measurements were localized on the standard brain using a well-established probabilistic registration procedure,²⁶ without adding any other imaging modalities. A hemispherical approximation of the brain was made to generate and quantify the topographic image on the complex brain surface, which simplified the three-dimensional (3-D) imaging to be two-dimensional (2-D) in a spherical coordinate system. Second, the test-retest reliability of the combined rTMS/fNIRS was assessed based on the brain-atlas guided images, by utilizing the experimental raw data, some of which were analyzed in our previous report.²⁰ The test-retest evaluation used three types of indices that had been introduced in the above-mentioned studies: (a) intraclass correlation coefficients that evaluated the reliability of $\Delta[\text{HbO}_2]$ at the group level; (b) individual reproducibility in sizes; and (c) degree of overlap of the activated regions or regions of interests (ROIs) that were identified in the topographic images.

2 Methods

2.1 Participants

Eleven healthy adults (nine males and two females, age range 20 to 49 years) were recruited from the local community. Participants could not have a past or current psychiatric disorder, history of a significant medical disorder, a presently unstable

medical condition, caffeinism (i.e., withdrawal symptoms with three days of abstinence), nicotine use, be pregnant or breast feeding, or currently taking any medication. The participants were screened with the Structured Clinical Interview for DSM-IV Axis I Disorders (SCID-I),²⁷ Transcranial Magnetic Stimulation Adult Safety Screen (TASS) form,²⁸ medical history review, and physical exam. A urine sample was obtained for a drug screen and a urine pregnancy test (if the participant was a woman with childbearing potential). Only eligible participants after the screening underwent simultaneous rTMS/fNIRS. The study protocol was approved by the University of Texas Southwestern Medical Center (Dallas, Texas) Institutional Review Board. Written informed consent was obtained from every participant prior to any study evaluations or procedures.

2.2 Experimental Procedures

The study involved two visits that were two to three days apart to reduce the chance of carryover effects from Visit 1 to Visit 2. In each visit, 1-Hz rTMS were applied over the participant's motor cortex and prefrontal cortex in two separate sessions using a Neuronetics Model 2100 CRS system (Malvern, Pennsylvania).⁹ A continuous-wave fNIRS system (CW5, TechEn Inc., Massachusetts)²⁹ was used to measure the cortical activities during rTMS. The CW5 system had magnetically compatible fibers and probe holder, which were inspected and confirmed in MRI right before the study.

Visit 1: Participant was positioned in the TMS device chair and adjusted to ensure comfort throughout the experiment. The primary motor cortex was investigated first because this region produced an observable and quantifiable behavioral change, which was used to adjust the stimulation intensity for each participant. An fNIRS probe was attached against the scalp with Velcro straps and elastic bandages. Inside the probe, the tips of the fibers were bent into an *L* shape so that the fibers could be pressed on the scalp. The fNIRS probe consisted of eight sources and 16 detectors, as shown in Fig. 1(a), which covered the primary motor cortices on both hemispheres. It provided a total of 28 channels (measurements) at a source-detector distance of 3.2 cm, 14 channels on each hemisphere. Once the fNIRS probe was in place, a TMS coil was placed over the left hemisphere and on top of the fNIRS probe, as shown in Fig. 1(b). The coil consisted of two loops of wire in a figure-of-eight arrangement; each loop was approximately 7 cm in diameter. The two loops were left-right oriented along a coronal line and adjusted to have approximately equal distance from the scalp. The location of maximal stimulation of the right abductor pollicis brevis (left motor cortex) was determined using the visual method.³⁰ The participant's individual RMT was determined using the T.M.S. Motor Threshold Assessment Tool³¹ three times and averaged. The stimulation intensity during the experiment was 120% RMT or 100% of TMS machine output if it could not reach 120% RMT.

Repetitive TMS was performed while the participant sat stably and awake. The stimulation epoch consisted of 10 s of 1-Hz magnetic pulses (i.e., 10 pulses) and 80 s of rest, which was repeated 15 times. The fNIRS data acquisition was initiated 1 min prior to the first epoch and ceased 1 min after the last epoch was completed, which resulted in a total recording time of 24 min and 30 s. After stimulation, the coordinates of coil position on the TMS device chair were recorded that would be used in Visit 2. The position of the fNIRS probe on the




Fig. 1 Schematics of simultaneous rTMS/fNIRS measurement on participant's head: (a) the geometry of the fNIRS probe that consisted of eight sources (filled circles) and 16 detectors (filled squares). The probe provided 28 nearest source-detector pairs (lines) in a distance of 3.2 cm, 14 pairs on each hemisphere; (b) the real placement of the fNIRS probe and TMS coil on a participant's motor cortex.

participant's head was determined by referring to the distance from the nasion to the center of the probe in a para-sagittal line.

The participant was given a short break of approximately 10 min after the stimulation session on the primary motor cortex. Then the fNIRS probe and the TMS coil were moved to the prefrontal cortex that was defined as 5 cm anterior to the motor cortex in a para-sagittal line. The stimulation was performed with the same intensity, frequency, and times as it was on the motor cortex, while the brain activity was recorded by fNIRS.

Visit 2: Participants returned for the second visit two to three days after Visit 1. Each of the participants was interviewed by a physician to ensure there were no changes to medical condition and to assess for any adverse effects related to the Visit 1 study procedures. Using the coordinates on the TMS device chair, the participant was placed back in the same position as in the first visit. The fNIRS probe was placed on the scalp at the same position of Visit 1 by referring to the distance from the nasion to the probe center in a para-sagittal line. The location of maximal stimulation of the abductor pollicis brevis was confirmed, and the RMT was determined again using the T.M.S. Motor Threshold Assessment Tool. This was to confirm that no significant change in RMT (greater than 10% change) occurred since Visit 1. The rTMS was performed with same intensity, frequency, and times as it was in Visit 1, on the primary motor cortex and the prefrontal cortex, respectively.

2.3 Data Screening and Processing

While some of the experimental data were reported in our previous publication,²⁰ there exist two major updates or differences in data analysis between the last study and this paper: (a) data

screening for qualified channels and subjects and (b) selection of approximate values of differential path-length factor (DPF) at two wavelengths. Our earlier study focused on channel-wise data analysis.²⁰ For this study, more rigorous selection criteria to screen the data were needed because a topographic image can be formed only when all measurement channels provide high-quality data.

In this study, the fNIRS data from each stimulation session were processed in the following steps:

First, the raw data were inspected for the entire time course to exclude: (a) some channels that had high-frequency instrumental noise larger than 15% of the baseline intensities and (b) some epochs that had significant discontinuities caused by facial muscle twitching.

Second, in order to derive a reliable image of the cortex based on all possibly good-quality channels available, we excluded stimulation sessions that had: (a) bad channels beneath the center of TMS coil as well as the contralateral counterparts; or (b) more than two bad channels on each hemisphere. After such data screening, the qualified subjects to be analyzed in this study were fewer than those in our previous report.²⁰

Third, a low-pass filter with a cut-off frequency at 0.2 Hz was applied to remove the high-frequency physiological noises in the qualified data, such as the arterial and respiratory waves.

Fourth, the channel-wise $\Delta[\text{HbO}_2]$ and $\Delta[\text{Hb}]$ were calculated based on the Modified Beer-Lambert Law,¹² in which we used $\text{DPF} = 6.8$ for laser wavelength at 690 nm and $\text{DPF} = 5.8$ for wavelength at 830 nm, according to the published data on the adult heads.³²

At last, the channel-wise $\Delta[\text{HbO}_2]$ and $\Delta[\text{Hb}]$ were averaged across all of the qualified epochs in the session to get epoch-averaged hemodynamic responses to rTMS.

2.4 Topographic Imaging

For each stimulation session, hemodynamic images of the cortical region under the fNIRS probe and the TMS coil were generated based on the epoch-averaged responses across all channels. To do so, first the fNIRS measurements were co-registered to a standard brain atlas in order to determine the cortical regions under the probe. Then a rapid 2-D topography approach with a spherical approximation of the brain surface was developed to form the hemodynamic images of the cortex.

Co-registration: A fundamental limitation of fNIRS is that it measures the hemodynamic changes from the scalp without any accurate knowledge of the underlying cortical structures. In other words, fNIRS as a stand-alone technology can provide only functional information without structural information. The ideal solution to this issue is acquiring each participant's brain MRI with fNIRS optodes attached, which, however, significantly reduces the convenience of fNIRS. As an alternative approach, probabilistic registration has been widely used in this field,²⁶ which registers the fNIRS measurements onto a standard brain template, instead of on each participant's own brain. In this approach, the optode positions on the participant's head are measured using a 3-D digitizer along with several cranial landmarks. The cranial landmarks are used to estimate the affine transformation from the real-world coordinate system to a target brain coordinate system.

The probabilistic registration was implemented in this study. To estimate the locations of fNIRS optodes over the motor cortex and the prefrontal cortex, three adult volunteers (two males and one female) were recruited afterwards. The fNIRS probe

was placed on their heads in the same way as that above-mentioned during simultaneous rTMS/fNIRS. The positions of optodes and five cranial landmarks (the nasion, inion, left and right earholes, and the vertex) were measured with a 3-D digitizer (PatriotTM, Polhemus, Colchester, Vermont, U.S.A.) and then loaded into a publically available toolbox (NIR-SPM, <http://bisp.kaist.ac.kr/NIRS-SPM.html>)³³ to elicit the positions of optodes and channels (defined as the midpoint of the source-detector pair) in the standard Montreal Neurological Institute (MNI) space.³⁴ The co-registration procedure was performed for each individual volunteer and then the results were averaged. The brain atlas we had used was ICBM 152 nonlinear, asymmetric template (ICBM 2009a Nonlinear Asymmetric $1 \times 1 \times 1$ mm template, <http://www.bic.mni.mcgill.ca/ServicesAtlases/ICBM152NLin2009>),^{35,36} This template covered a broad range of age (18.5 to 43.5 years), and the anatomical structures were segmented using ANIMAL + INSECT algorithm.³⁷ Figure 2 shows the cortical surface of the brain extracted from the ICBM 152 template.

Topography: Given the sparse probe geometry and the single source-detector distance used in this study, the current fNIRS data was expected to have low specificity in depth. Therefore, topographic images of the cortical surface in the stimulated regions were generated by interpolating the epoch-averaged, channel-wise data. With limitations in depth specificity, the topography approach provides us with a fast and efficient approximation to obtain hemodynamic images of the cortical region in response to the combined rTMS/fNIRS. It avoids the complex computation to simulate the light propagation in heterogeneous tissues and to solve the inverse problem that would be encountered in optical tomography.

2.5 Formation of “EasyTopo”

Brain atlas-based topography has been used in a few published reports.^{33,38} In general, the data is interpolated on a stereotaxic convex defined by a cohort of 2-D surfaces, which indeed does not match the brain surface perfectly. Sometimes the data have to be further extrapolated. In this study, we developed a rapid topography for brain imaging (referred to as “EasyTopo”) by




Fig. 2 Brain surface for optical topography with two coordinates systems. The brain surface was extracted from the standard ICBM 152 template in stereotaxic coordinates (x, y, z) . In optical topography, the stereotaxic coordinates were converted into spherical coordinates (r, θ, φ) based on a hemispherical approximation of the brain surface, where r is radial distance, θ is azimuthal angle, and φ is polar angle.

implementing angular interpolation in a spherical coordinate system. EasyTopo was derived based on the fact that the cortical layer of the brain is approximately a hemispherical surface. Therefore the stereotaxic coordinates of the brain surface in (x, y, z) were converted into spherical coordinates in (r, θ, φ) (refer to Fig. 2). As schematized in Fig. 3(a) to 3(c), the brain activation in the spherical coordinate system represented a distribution mainly in 2-D (θ, φ) space. Therefore, 2-D angular interpolation was conducted in the (θ, φ) space to form an image of the brain activation, which was then projected back onto the brain surface in original 3-D stereotaxic coordinates.

It is noted that EasyTopo does not choose a constant radius. Each surface pixel is marked by its own radius, r_i , as well as its azimuthal and polar angles (θ_i and φ_i). However, r_i is not reflected in 2-D (θ, φ) map. Analogy to this procedure is how the earth is drawn in a latitude-longitude map: each local place (i.e., a surface pixel) has its elevation (i.e., radius) as well as its latitude and longitude degrees, but only the latitude and longitude degrees are used for mapping. In other words, we can compress all of the surface pixels of the brain onto a 2-D (θ, φ) map even though their radii or elevations differ. In EasyTopo, the hemodynamic data are interpolated on such a map [see Fig. 3(b)], which contains all of the pixels from the brain surface with respective radius values, not a uniform or equal radius. After all the surface pixels with interpolated values are projected back onto the brain according to their radii, the eventual topographic image [see Fig. 3(c)] is continuous. In addition, we wish to point out a special case: it is possible for a pixel on the gyrus and a pixel on the sulci to get the same projected value in 2-D (θ, φ) space if they share the same latitude and longitude degrees (i.e., the sulci pixel is beneath the gyrus pixel, and they are located along the same radial direction). In this case, we should ignore the sulci pixel since fNIRS is much more sensitive to the shallower tissues.

Compared with the previous topography methods,^{33,38} EasyTopo was more computationally efficient and did not require any data interpolation/extrapolation in 3-D stereotaxic space. Another advantage of EasyTopo is that the data between two adjacent channels is interpolated along their included angle in the spherical coordinate system, rather than along a straight line going under the brain surface. The former interpolation is more physiologically meaningful, agrees better with the reality, and serves more accurate in image representation.

2.6 Identifying Regions of Interest

To characterize the spatial pattern of rTMS-evoked response, a k -means clustering algorithm was used to identify regions of interest (ROIs) from the background. This method had been used in our previous study³⁹ to identify positive motor response from the background. In this study, the rTMS-evoked response could be either positive or negative. Therefore the algorithm was modified, and it segmented the topographic image into two to three regions through one of the following cases:

- (1) The rTMS-evoked hemodynamic response, H , was mainly negative; there was no or negligible positive response, as expressed by a mathematical criterion of $|H_{\min}|/2 > H_{\max}$, where H_{\min} and H_{\max} represent the negative minimum and positive maximum of hemodynamic response. In this case, two initial seeds (i.e., H_{\min} and zero) were used to result in two




Fig. 3 Schematics of 2-D optical topography in a spherical coordinate system: (a) the (θ, φ) lattices on the brain surface defined in the spherical coordinate system. As the brain surface is approximately in a hemispherical shape, the cortical activation is a distribution mainly on the (θ, φ) lattices regardless the radius; (b) the topographic image of motor cortex generated through data interpolation in (θ, φ) space. The solid curves in the image outline the ROIs identified by k -means clustering; (c) the topographic image projected back from (θ, φ) space to the original stereotaxic coordinate system and overlaid on the brain surface.

clusters of pixels: ROIs initialized by H_{\min} , and background initialized by zero.

- (2) The rTMS-evoked hemodynamic response, H , was mainly negative, but having nonnegligible positive response, as expressed by $|H_{\min}|/2 \leq H_{\max}$. In this case, three initial seeds (i.e., H_{\min} , zero, and H_{\max}) were used to result in three clusters of pixels: ROIs, background, and small positive response regions, respectively.
- (3) The rTMS-evoked hemodynamic response, H , was mainly positive; there was no or negligible negative response, as expressed by $H_{\max}/2 > |H_{\min}|$. In this case, two initial seeds (i.e., H_{\max} and zero) were used to give rise to two clusters of pixels: ROIs and background, respectively.
- (4) The rTMS-evoked hemodynamic response, H , was mainly positive, with nonnegligible negative response, as expressed by $H_{\max}/2 \leq |H_{\min}|$. In this case, three initial seeds (i.e., H_{\max} , zero, and H_{\min}) were used to lead to three clusters of pixels: ROIs, background, and small negative regions, respectively.

For convenience, the clustering algorithm was applied on the interpolated images in (θ, φ) space rather than the eventual images in stereotaxic coordinates, as schematized in Fig. 3(b). Once the ROIs were identified, the solid angle of each pixel in the ROIs was computed as:

$$\Omega_i = \cos \varphi_i d\varphi d\theta, \quad (1)$$

where Ω_i and φ_i denote the solid angle and the polar angle of the i 'th pixel in the identified ROIs, respectively; $d\varphi$ and $d\theta$ denote the step of the polar angle and the step of the azimuthal angle, respectively. Then the total area of the identified ROIs, A_{ROIs} , could be estimated as:

$$A_{\text{ROIs}} = \sum_{i \in \text{ROIs}} \Omega_i r_i^2, \quad (2)$$

where r_i denotes the radius of the i 'th pixel in the identified ROIs, which was also obtained via interpolation in (θ, φ) space.

In this study, the A_{ROIs} was estimated in the hemisphere ipsilateral to the rTMS (i.e., left hemisphere) and in the one contralateral to the rTMS (i.e., right hemisphere) separately. To evaluate the laterality of A_{ROIs} between two hemispheres, we introduced a lateralization factor, L , which was computed as:⁴⁰

$$L = \frac{A_{\text{ROIs_ipsi}} - A_{\text{ROIs_contra}}}{A_{\text{ROIs_ipsi}} + A_{\text{ROIs_contra}}}, \quad (3)$$

where $A_{\text{ROIs_ipsi}}$ and $A_{\text{ROIs_contra}}$ denote the areas of ROIs in the ipsilateral hemisphere and in the contralateral hemisphere, respectively. An L value of 1 reflects a complete ipsilateral response, an L value of -1 reflects a complete contralateral response, and an L value around 0 reflects a bilateral response.

2.7 Assessing Test-Retest Reliability

To comprehensively evaluate the test-retest reliability of the rTMS-evoked hemodynamic response, we used three types of indices that have been introduced in Ref. 25: (a) the intraclass correlation coefficients, ICCs; (b) reproducibility of the identified ROIs in sizes, R_{size} , in two repeated sessions; and (c) degree of overlap, R_{overlap} , between the identified ROIs in two repeated sessions.

The first index, ICC, assesses the reliability of the quantified hemodynamic changes in two repeated sessions based on a one-way random effect model for consistency measurements.^{22,24,25,41} In this study, the hemodynamic changes within the identified ROIs of each session were averaged as a measure of rTMS-evoked response in amplitude. Two ICCs were defined to evaluate the reliability of a single measurement, $\text{ICC}_{\text{SINGLE}}$, and the reliability of the mean of multiple measurements, $\text{ICC}_{\text{AVERAGE}}$:

$$\text{ICC}_{\text{SINGLE}} = \frac{\text{BMS} - \text{WMS}}{\text{BMS} + (k - 1)\text{WMS}} \quad (4)$$

and

$$\text{ICC}_{\text{AVERAGE}} = \frac{\text{BMS} - \text{WMS}}{\text{BMS}}, \quad (5)$$

where BMS is the between-subject mean square, WMS is the within-subject mean square, and k is the number of measurements (in this study $k = 2$).

The second index, R_{size} , assesses the agreement of the identified ROIs in size between two repeated sessions⁴² and was defined as:

$$R_{\text{size}} = 1 - \frac{|\text{ROI}_1 - \text{ROI}_2|}{\text{ROI}_1 + \text{ROI}_2}, \quad (6)$$

where ROI_1 and ROI_2 denote the sizes of respective ROIs in two repeated sessions.

The third index, R_{overlap} , assesses the degree of overlap of the identified ROIs between two repeated sessions and was defined as:

$$R_{\text{overlap}} = \frac{2\text{ROI}_{\text{overlap}}}{\text{ROI}_1 + \text{ROI}_2}, \quad (7)$$

where $\text{ROI}_{\text{overlap}}$ denotes the overlapped area between the activated ROIs in two repeated sessions.

For all of the indices described above (i.e., $\text{ICC}_{\text{SINGLE}}$, $\text{ICC}_{\text{AVERAGE}}$, R_{size} , and $\text{ROI}_{\text{overlap}}$), values of ≥ 0.80 were considered as highly reliable, ≥ 0.60 as moderately reliable, and < 0.60 as weakly reliable, according to the criteria given by Charter.⁴³

3 Results

3.1 Quantification and Topography of rTMS-Evoked Hemodynamic Response

Among all of the 11 participants, three participants had consistently noisy fNIRS data across all the channels in motor stimulation sessions. It was presumably due to dense and/or dark-colored hair, which could lead to loose contact of fNIRS optodes on the head or high light absorption⁴⁴ if some hair was left between the optode and scalp. Thus these three participants were excluded from data analysis (i.e., $N = 8$ for motor stimulation sessions). For the same reason, we excluded one participant when rTMS was on the prefrontal cortex (i.e., $N = 10$ for prefrontal stimulation sessions).

For all of the remaining participants with qualified data, we further excluded a few channels with high-frequency instrumental noise above the threshold. Those excluded channels were approximately 5.8% of total channels in motor stimulation sessions and 7.3% of total channels in prefrontal stimulation sessions. Also, the excluded channels were not directly beneath the center of TMS coil or its contralateral counterparts, which likely

had the maximum response to the stimulation. Given the diffusion nature of fNIRS, the response signal at every excluded channel was approximately replaced with an averaged signal from its nearest neighboring channels. This procedure ensured the topographic image to be relatively continuous across the region of a bad channel. Note that the location of a channel was defined as the midpoint between its nearest source and detector. The nearest neighboring channels of an excluded channel was those that have the minimal Euclidean distance to the excluded channel.

For the qualified data after screening, previously we have reported a deactivation pattern at the group level, indicated by robust decreases in $\Delta[\text{HbO}_2]$ and slight increases in $\Delta[\text{Hb}]$, during and shortly after rTMS.²⁰ As an example, Fig. 4(a) and 4(b) shows the grand-averaged temporal responses seen in Visit 1 by selecting two to four channels under the center of TMS coil (i.e., on the ipsilateral side), at the motor cortex and the prefrontal cortex, respectively. At both the motor cortex and the prefrontal cortex, the most significant changes are seen during the first 20 s of the epoch. Therefore, the channel-wise data in every session was averaged across the first 20 s to form the topographic images.

It is noted that the curves in both Fig. 4(a) and 4(b) look slightly different from those seen in our previous report (Fig. 2 in Ref. 20). Two reasons cause the difference: (a) In this study, eight subjects in the motor stimulations and 10 subjects in the prefrontal stimulation sessions had been selected to form topographic images of respective cortices. The number of qualified subjects in Ref. 20 was 11 for both motor and prefrontal stimulation studies, based on channel-wise analysis. (b) In this study, we used $\text{DPF} = 6.8$ for 690 nm and $\text{DPF} = 5.8$ for 830 nm in computation of hemoglobin concentrations. These values were estimated based on the previous publication¹² to approximate light scattering effects in the human brain. In Ref. 20, we had used $\text{DPF} = 1$ for both 690 and 830 nm, which significantly deviated from the real DPF values in human brain tissues. The current method corrects or improves the accuracy of quantification greatly with respect to the previous calculation.

Figure 5(a) to 5(d) shows the grand-averaged ($N = 8$) topographic images at the motor cortex, obtained in Visit 1. In the ipsilateral hemisphere, the deactivation pattern is localized under the TMS coil, which is indicated by distinct $\Delta[\text{HbO}_2]$ decrease around the motor cortex strip. The deactivation pattern in the contralateral hemisphere is more diffused. Figure 6(a) to 6(d) shows the grand-averaged ($N = 10$) topographic images




Fig. 4 Grand-averaged temporal brain responses to rTMS (solid curve: mean; shaded region: standard deviation) at: (a) the primary motor cortex ($N = 8$, Visit 1); (b) the prefrontal cortex ($N = 10$, Visit 1). At each cortical region, the hemodynamic responses (red: $\Delta[\text{HbO}_2]$; blue: $\Delta[\text{Hb}]$) were attained by selecting two to four channels under the center of TMS coil for each individual participant. The vertical lines (dark green) in each plot indicate the rTMS pulses, which were applied in the first 10s of the epoch.




Fig. 5 Grand-averaged topographic images ($N = 8$, Visit 1) at the motor cortex: (a) the $\Delta[\text{HbO}_2]$ image in the ipsilateral hemisphere; (b) the $\Delta[\text{Hb}]$ image in the ipsilateral hemisphere; (c) the $\Delta[\text{HbO}_2]$ image in the contralateral hemisphere; (d) the $\Delta[\text{Hb}]$ image in the contralateral hemisphere. Each image was normalized by its minima [(a) and (c)] or maxima [(b) and (d)]. The location and orientation of TMS coil over the motor cortex is indicated in [(a) and (b)].




Fig. 6 Grand-averaged topographic images ($N = 10$, Visit 1) at the prefrontal cortex: (a) the $\Delta[\text{HbO}_2]$ image in the ipsilateral hemisphere; (b) the $\Delta[\text{Hb}]$ image in the ipsilateral hemisphere; (c) the $\Delta[\text{HbO}_2]$ image in the contralateral hemisphere; (d) the $\Delta[\text{Hb}]$ image in the contralateral hemisphere. Each image was normalized by its minima [(a) and (c)] or maxima [(b) and (d)]. The location and orientation of TMS coil over the prefrontal cortex is indicated in [(a) and (b)].

at the prefrontal cortex, again from Visit 1. Similarly, the deactivation pattern in the ipsilateral hemisphere is more localized.

3.2 Assessment of Test-Retest Reliability for rTMS/fNIRS

The test-retest reproducibility was evaluated based on the $\Delta[\text{HbO}_2]$ images only because it demonstrated robust changes associated with rTMS. We neglected the $\Delta[\text{Hb}]$ images since the $\Delta[\text{Hb}]$ changes were insignificant at the group level. Most participants had mainly decreased or negative $\Delta[\text{HbO}_2]$ images.

Table 1 Area of ROIs in the ipsilateral hemisphere, $A_{\text{ROIs_ipsi}}$ (cm^2), and in the contralateral hemisphere, $A_{\text{ROIs_contra}}$ (cm^2), as well as laterality between two hemispheres, L , quantified after k -means clustering.

	Primary motor cortex		Prefrontal cortex	
	Visit 1	Visit 2	Visit 1	Visit 2
$A_{\text{ROIs_ipsi}}$ (cm^2) ^a	6.9 ± 5.7	11.5 ± 4.6	16.6 ± 5.9	16.3 ± 7.1
$A_{\text{ROIs_contra}}$ (cm^2) ^b	6.1 ± 7.5	8.8 ± 5.2	12.5 ± 4.9	12.1 ± 6.2
L	0.19 ± 0.84	0.22 ± 0.38	0.33 ± 0.40	0.40 ± 0.14

^a $p < 0.01$ (two-sample t-test) between the primary motor cortex and the prefrontal cortex.

^b $p < 0.02$ (two-sample t-test) between the primary motor cortex and the prefrontal cortex.

Consequently, in k -means clustering, the cluster initialized by the minimum value of the image represented the ROIs that were significantly deactivated by rTMS. Interestingly, we did observe one participant in two repeated motor stimulation sessions and another participant in two repeated prefrontal stimulation sessions with consistently increased or positive $\Delta[\text{HbO}_2]$ images. In these special cases, the cluster initialized by the maximum value of the image represented the ROIs. Table 1 summarizes the resultant areas of ROIs in each hemisphere and the laterality between two hemispheres after k -mean clustering. At the primary motor cortex, the rTMS evoked a grand-averaged area of 6.1 to 11.5 cm^2 in each hemisphere. At the prefrontal cortex, the rTMS evoked a grand-averaged area of 12.1 to 16.6 cm^2 in each hemisphere, which is significantly larger than that at the primary motor cortex (t-test, $p < 0.01$ for the ipsilateral hemisphere and $p < 0.02$ for the contralateral hemisphere). At both the primary motor cortex and the prefrontal cortex, the lateralization factor indicates that the ROIs in the ipsilateral hemisphere are slightly larger than those in the contralateral hemisphere.

At the primary motor cortex, high reliability for single measure of $\Delta[\text{HbO}_2]$ (mean value within the identified ROIs) with $\text{ICC}_{\text{SINGLE}} = 0.86$ and high reliability for average measure of $\Delta[\text{HbO}_2]$ with $\text{ICC}_{\text{AVERAGE}} = 0.92$ were attained at the group level ($N = 8$). However, it is noteworthy that the positive $\Delta[\text{HbO}_2]$ response from one participant greatly enlarged the between-subject mean square (BMS) when calculating $\text{ICC}_{\text{SINGLE}}$ and $\text{ICC}_{\text{AVERAGE}}$. If we excluded this participant from data analysis, moderate reliability for single measure of $\Delta[\text{HbO}_2]$ with $\text{ICC}_{\text{SINGLE}} = 0.68$ and high reliability for average measure of $\Delta[\text{HbO}_2]$ with $\text{ICC}_{\text{AVERAGE}} = 0.81$ were attained ($N = 7$). At the individual level, the identified ROIs showed averagely high agreement in size ($R_{\text{size}} = 0.81 \pm 0.11$) and nearly moderate degree of overlap ($R_{\text{overlap}} = 0.56 \pm 0.16$). The individual results of mean $\Delta[\text{HbO}_2]$ within the identified ROIs, R_{size} and R_{overlap} are summarized in Table 2.

Similarly, at the prefrontal cortex, high reliability for single measure of $\Delta[\text{HbO}_2]$ with $\text{ICC}_{\text{SINGLE}} = 0.81$ and high reliability for average measure of $\Delta[\text{HbO}_2]$ with $\text{ICC}_{\text{AVERAGE}} = 0.89$ were attained at the group level ($N = 10$). Here it is also noted that the positive $\Delta[\text{HbO}_2]$ response from one participant greatly enlarged BMS. If we excluded this participant from data analysis, moderate reliability for single measure of $\Delta[\text{HbO}_2]$ with

Table 2 Individual results of mean $\Delta[\text{HbO}_2]$ (μM) within the identified ROIs, R_{size} and R_{overlap} quantified in the motor stimulation sessions.

Participant #	$\Delta[\text{HbO}_2]$ (μM)		R_{size}	R_{overlap}
	Visit 1	Visit 2		
1	-1.25	-0.70	0.74	0.54
2	-0.30	-0.20	0.75	0.37
3*	1.19	0.65	0.95	0.72
7	-0.16	-0.21	0.89	0.54
8	-0.28	-0.25	0.82	0.69
9	-0.52	-0.45	0.94	0.75
11	-0.70	-1.04	0.61	0.33
12	-0.28	-0.55	0.80	0.56

*Subject who had consistent positive response ($\Delta[\text{HbO}_2] > 0$) to rTMS in two visits.

Table 3 Individual results of mean $\Delta[\text{HbO}_2]$ (μM) within the identified ROIs, R_{size} and R_{overlap} quantified in the prefrontal stimulation sessions.

Participant #	$\Delta[\text{HbO}_2]$ (μM)		R_{size}	R_{overlap}
	Visit 1	Visit 2		
1	-1.42	-0.68	0.92	0.54
2	-0.64	-0.55	0.93	0.61
4*	0.41	0.37	0.87	0.72
5	-0.54	-0.61	0.90	0.82
6	-0.88	-1.17	0.98	0.93
7	-0.18	-0.16	0.70	0.38
8	-0.40	-0.41	0.81	0.59
9	-0.86	-0.66	0.97	0.76
11	-0.55	-0.37	0.95	0.63
12	-0.41	-0.58	0.78	0.72

*Subject who had consistent positive response ($\Delta[\text{HbO}_2] > 0$) to rTMS in two visits.

$\text{ICC}_{\text{SINGLE}} = 0.60$ and moderate reliability for average measure of $\Delta[\text{HbO}_2]$ with $\text{ICC}_{\text{AVERAGE}} = 0.75$ were attained ($N = 9$). At the individual level, the identified ROIs showed averagely high agreement in size ($R_{\text{size}} = 0.88 \pm 0.09$) and moderate degree of overlap ($R_{\text{overlap}} = 0.67 \pm 0.15$). The individual results of mean $\Delta[\text{HbO}_2]$ within the identified ROIs, R_{size} and R_{overlap} are summarized in Table 3.

4 Discussions

Several studies,¹⁴⁻¹⁹ including our previous one,²⁰ have demonstrated the great potential of fNIRS to objectively assess the

neural effects of rTMS at cortical regions beneath and distant (e.g., the contralateral counterparts) from the coil. Because the optical signal in fNIRS does not interact with the magnetic or electric fields produced by the TMS coil, the combination of two technologies is relatively simple. In our experimental setup, the fNIRS probe was placed between the scalp and the coil, which kept the TMS coil about 1.5 cm away from the scalp. As the magnetic field falls off rapidly in this additional distance, the participants' RMTs were greater than they otherwise would have been without the fNIRS probe. This shortcoming can be minimized by using compact optodes. Another issue we have observed is that rTMS sometimes made facial muscles twitch, which introduced transient artifacts into fNIRS data. In this study, we excluded all of the epochs with visible artifacts, and the remaining epochs were still sufficient to deduce reliable epoch-averaged responses. However, it is noteworthy that the artifacts in fNIRS data can be removed in a more efficient way without excluding the relevant epochs, as demonstrated in a recent study.⁴⁵

The new features in this paper distinguished from our previous one²⁰ are two: one is to investigate spatial distributions of rTMS-evoked hemodynamic responses, and the other is to assess the combined rTMS/fNIRS measurement reproducibility. We discuss these two aspects in the following two subsections.

4.1 Spatial Characterization of rTMS-Evoked Response

A rapid optical topography on the basis of a standard brain atlas was developed to spatially characterize the rTMS-evoked responses on both ipsilateral and contralateral hemispheres at the primary motor cortex and the prefrontal cortex, respectively. Spatial pattern of the topographic images, represented by ROIs, was identified with a k -means clustering algorithm. These methods enabled objective determination of the rTMS-evoked response measured with fNIRS, and quantitative evaluations of the test-retest reliability of this combined technology at the individual level. Interestingly, we also found that the rTMS-evoked responses had greater (nearly two times) hemodynamic changes (see Fig. 4) and larger area of ROIs (see Table 1) at the prefrontal cortex as compared with those at the primary motor cortex. The difference in the areas of ROIs was statistically significant. These differences probably result from the differences in neurophysiology or neuro-anatomy of the different brain regions in response to rTMS. One possible explanation is as follows: the intensity of magnetic pulses is attenuated by the distance from the coil; the optical signals are also attenuated greatly by the scalp and the skull before reaching the brain. Therefore, a shorter scalp-to-brain distance could result in (a) a larger area of magnetic stimulation that generates the electrical potential above the firing threshold of the neurons as well as (b) a higher measurement sensitivity of fNIRS. Unfortunately, the sparsity of our current fNIRS probe did not support reliable 3-D tomographic imaging to verify this explanation; it needs to be further investigated or confirmed in future. Nevertheless the findings that the rTMS-evoked responses had greater $[\text{HbO}_2]$ changes and larger extended ROIs in the prefrontal cortex than in the primary motor cortex are new, providing valuable evidence and observation to neuroscientists and psychiatrists for a better understanding of neurological effects of rTMS on the human brain.

4.2 Test-Retest Assessment of rTMS/fNIRS

For the second aspect of this paper, namely, the test-retest assessment, it is noteworthy that several measurement conditions in this study are essential: first, the participant was placed in the same position by referring to the coordinates on the TMS device chair; second, the fNIRS probe was placed in the same position by referring to the distance from the nasion to the center of the probe; third, the individual RMT was determined in both visits and confirmed unchanged. The intensity of rTMS was identical between the two visits. The overall lessons learned in test-retest reliability of simultaneous rTMS/fNIRS are as follows: At the primary motor cortex, moderate-to-high test-retest reliability between two repeated sessions could be achieved in quantity of $\Delta[\text{HbO}_2]$, represented by $\text{ICC}_{\text{SINGLE}}$ and $\text{ICC}_{\text{AVERAGE}}$. The identified ROIs showed overall high reproducibility in size and nearly moderate degree of overlap between two repeated sessions (see Table 2). At the prefrontal cortex, the quantity of $\Delta[\text{HbO}_2]$ showed moderate test-retest reliability between two repeated sessions. The identified ROIs showed overall high reproducibility in size and moderate degree of overlap between two repeated sessions (see Table 3).

Several earlier studies given by Refs. 22 to 24 focused on the fNIRS reliability with the test-retest interval between three weeks and one year. The purpose of this study was to test the stability of the rTMS/fNIRS measurement method, not to test the stability of the brain. The acute effects of a single session of rTMS are generally thought to last about 20 to 30 min with standard pulse sequences as used in this protocol.⁴⁶ The time interval of two to three days used in this study was chosen for the exact purpose of reducing any chance of a carryover effect while minimizing possible physiologic changes in the brain that could account for a difference. Even if there was a carryover effect from the rTMS, it would be expected to reduce the reliability of the technique and not improve it. The limitation of such a protocol design is that it does not test the variability over a wide range of time intervals; further work is planned in this direction.

Further exploration of the test-retest reliability of combined rTMS/fNIRS can be neurological, technical, or both. Neurologically, any cumulative brain effects across the two visits could certainly affect the test-retest reliability, which had been a concern in long-term daily rTMS treatments.⁴⁶ However, cumulative effects were very unlikely to occur in this study because the stimulation sessions were short (150 s of 1-Hz stimulation per session), and the two visits were two to three days apart. Technically, the quality of fNIRS data can be affected by the individual brain anatomy, spatial geometry, and placement-variability of the fNIRS probes.^{22–25} In particular, two factors need to be considered when referring to the results in the present paper: first, although the location of motor cortex was determined based on the maximal stimulation of the right abductor pollicis brevis, the location of prefrontal cortex was defined as 5 cm anterior to the motor cortex in a para-sagittal line. The exact same prefrontal region could not be guaranteed to have been stimulated across all subjects, depending on the cranial size. Future studies should consider using other orientation points, such as markers from the 10 to 20 international EEG system, to determine the location of prefrontal regions more precisely. Second, because of the initial experimental design, the locations of the fNIRS probe on each participant's head were not measured for co-registration with a standard brain atlas during simultaneous rTMS/fNIRS. To estimate the cortical

regions under the fNIRS probe, digitizer readings were obtained from other three volunteers who did not undergo simultaneous rTMS/fNIRS. Their co-registration results were averaged to provide a rough, probabilistic estimation about the cortical location interrogated by the probe in order to implement the brain atlas-based topography. However, the head shape and anatomical structures vary across participants, so nonsubject-specific co-registration may introduce location errors. For future studies, brain topography should be performed according to each participant's own optode-brain relationship in order to minimize the location-variability.

The practice of replacing excluded channels with an averaged signal from its nearest neighboring channels was an approximate step to avoid a discontinuous or void area in topographic images. Since each channel in the probe array (Fig. 1) corresponded to a certain spatial region in the tomographic image, the averaged signal from nearest neighboring channels (corresponding to other regions) was only an estimation to reflect or replace the missing signal at the excluded channel. Such a practice would introduce spatial errors at the excluded channels for the identified spatial distribution of rTMS-evoked hemodynamic response and ROIs, thus leading to errors in the quantified R_{size} and R_{overlap} . The above statement implies that the more measurement channels were excluded for data analysis, the less reliable the test-retest results could be. Therefore it is critical and important in the experimental setup to improve optode-scalp contact and thus to reduce the number of excluded channels for data analysis. In this study, while we did have several measurement channels excluded due to highly noisy signals, they were only 5.8% of total channels in motor stimulation sessions and 7.3% of total channels in prefrontal stimulation sessions, also not directly beneath the center of TMS coil or its contralateral counterparts. Therefore the test-retest assessment of combined rTMS/fNIRS on the motor and prefrontal cortices should be reasonably reliable, as reported in Sec. 3.

Finally, the hemodynamic responses presented in this paper are only valid for a specific rTMS paradigm. TMS activates a mixture of neuron populations with different electrical thresholds; its eventual outcomes depend on stimulation intensity, frequency, shape and orientation of the coil, and other factors. For example, several previous studies have suggested that stimulation at a higher frequency than 1 Hz tends to increase cortical excitability, whereas stimulation at a lower frequency (≤ 1 Hz) tends to decrease cortical excitability.^{47,48} In this study, cortical deactivations associated with 1-Hz rTMS were identified among most participants. However, we did observe that one participant during motor stimulation and another participant during prefrontal stimulation had reproducible cortical activations. These special cases reveal and imply the complexity of rTMS-evoked brain responses, especially when the participant pool becomes large. In our future studies, we plan to use other stimulation frequencies besides 1 Hz to examine frequency dependence of the rTMS-evoked hemodynamic responses and thus to gain a better understanding of the brain under rTMS treatment.

Acknowledgments

The authors thank Mr. Sameer Dhamne for his assistance in data collection. Funding was provided by a contribution from the Jack A. Burke Estate, Grant-in-kind of supplies and TMS machine from Neuronetics, supplies and fNIRS machine from UT Arlington, and Grant-in-kind of supplies and personnel from UT Southwestern Medical Center.

References

1. M. S. George and R. H. Belmaker, *Transcranial Magnetic Stimulation in Clinical Psychiatry*, American Psychiatric Publishing, Inc. Washington D.C. (2007).
2. M. J. Edwards, P. Talelli, and J. C. Rothwell, "Clinical applications of transcranial magnetic stimulation in patients with movement disorders," *Lancet Neurol.* **7**(9), 827–840 (2008).
3. T. Burt, S. H. Lisanby, and H. A. Sackeim, "Neuropsychiatric applications of transcranial magnetic stimulation: a meta-analysis," *Int. J. Neuropsychopharmacol.* **5**(1), 73–103 (2002).
4. J. L. Couturier, "Efficacy of rapid-rate repetitive transcranial magnetic stimulation in the treatment of depression: a systematic review and meta-analysis," *J. Psychiatry. Neurosci.* **30**(2), 83–90 (2005).
5. P. E. Holtzheimer, III, J. Russo, and D. H. Avery, "A meta-analysis of repetitive transcranial magnetic stimulation in the treatment of depression," *Psychopharmacol. Bull.* **35**(4), 149–169 (2001).
6. F. A. Kozel and M. S. George, "Meta-analysis of left prefrontal repetitive transcranial magnetic stimulation (rTMS) to treat depression," *J. Psychiatr. Pract.* **8**(5), 270–275 (2002).
7. J. L. Martin et al., "Repetitive transcranial magnetic stimulation for the treatment of depression. Systematic review and meta-analysis," *Br. J. Psychiatry* **182**, 480–491 (2003).
8. M. S. George et al., "Daily left prefrontal transcranial magnetic stimulation therapy for major depressive disorder: a sham-controlled randomized trial," *Arch Gen Psychiatry* **67**(5), 507–516 (2010).
9. J. P. O'Reardon et al., "Efficacy and safety of transcranial magnetic stimulation in the acute treatment of major depression: a multisite randomized controlled trial," *Biol. Psychiatry* **62**(11), 1208–1216 (2007).
10. A. Villringer and B. Chance, "Non-invasive optical spectroscopy and imaging of human brain function," *Trends Neurosci.* **20**(10), 435–442 (1997).
11. D. A. Boas, A. M. Dale, and M. A. Franceschini, "Diffuse optical imaging of brain activation: approaches to optimizing image sensitivity, resolution, and accuracy," *Neuroimage* **23**(Suppl. 1), S275–S288 (2004).
12. M. Cope et al., "Methods of quantitating cerebral near infrared spectroscopy data," *Adv. Exp. Med. Biol.* **222**, 183–189 (1988).
13. E. Amaro, Jr. and G. J. Barker, "Study design in fMRI: basic principles," *Brain Cogn.* **60**(3), 220–232 (2006).
14. Y. Noguchi, E. Watanabe, and K. L. Sakai, "An event-related optical topography study of cortical activation induced by single-pulse transcranial magnetic stimulation," *Neuroimage* **19**(1), 156–162 (2003).
15. Y. Hada et al., "Detection of cerebral blood flow changes during repetitive transcranial magnetic stimulation by recording hemoglobin in the brain cortex, just beneath the stimulation coil, with near-infrared spectroscopy," *Neuroimage* **32**(3), 1226–1230 (2006).
16. N. Hanaoka et al., "Deactivation and activation of left frontal lobe during and after low-frequency repetitive transcranial magnetic stimulation over right prefrontal cortex: a near-infrared spectroscopy study," *Neurosci. Lett.* **414**(2), 99–104 (2007).
17. H. Mochizuki et al., "Cortical hemoglobin-concentration changes under the coil induced by single-pulse TMS in humans: a simultaneous recording with near-infrared spectroscopy," *Exp. Brain Res.* **169**(3), 302–310 (2006).
18. H. Shibasaki, "Human brain mapping: hemodynamic response and electrophysiology," *Clin. Neurophysiol.* **119**(4), 731–743 (2008).
19. Y. Aoyama et al., "Stimulus intensity dependence of cerebral blood volume changes in left frontal lobe by low-frequency rTMS to right frontal lobe: a near-infrared spectroscopy study," *Neurosci. Res.* **63**(1), 47–51 (2009).
20. F. A. Kozel et al., "Using simultaneous repetitive transcranial magnetic stimulation/functional near infrared spectroscopy (rTMS/fNIRS) to measure brain activation and connectivity," *Neuroimage* **47**(4), 1177–1184 (2009).
21. F. A. Kozel and M. H. Trivedi, "Developing a neuropsychiatric functional brain imaging test," *Neurocase* **14**(1), 54–58 (2008).
22. A. Watanabe et al., "Cerebrovascular response to cognitive tasks and hyperventilation measured by multichannel near-infrared spectroscopy," *J. Neuropsychiatry Clin. Neurosci.* **15**(4), 442–449 (2003).
23. M. M. Plichta et al., "Event-related functional near-infrared spectroscopy (fNIRS): are the measurements reliable?" *Neuroimage* **31**(1), 116–124 (2006).
24. M. M. Plichta et al., "Event-related functional near-infrared spectroscopy (fNIRS) based on craniocerebral correlations: Reproducibility of activation?" *Hum. Brain Mapp.* **28**(8), 733–741 (2007).
25. H. Zhang et al., "Test-retest assessment of independent component analysis-derived resting-state functional connectivity based on functional near-infrared spectroscopy," *Neuroimage* **55**(2), 607–615 (2011).
26. A. K. Singh et al., "Spatial registration of multichannel multi-subject fNIRS data to MNI space without MRI," *Neuroimage* **27**(4), 842–851 (2005).
27. M. B. First et al., *Structured Clinical Interview for DSM-IV Axis I Disorders, Clinician Version (SCID-CV)*, American Psychiatric Press, Inc., Washington, DC (1996).
28. J. C. Keel, M. J. Smith, and E. M. Wassermann, "A safety screening questionnaire for transcranial magnetic stimulation," *Clin. Neurophysiol.* **112**(4), 720 (2001).
29. M. A. Franceschini et al., "Diffuse optical imaging of the whole head," *J. Biomed. Opt.* **11**(5), 054007 (2006).
30. S. Pridmore et al., "Motor threshold in transcranial magnetic stimulation: a comparison of a neurophysiological and a visualization of movement method," *J. ECT* **14**(1), 25–27 (1998).
31. J. J. Borckardt et al., "Estimating resting motor thresholds in transcranial magnetic stimulation research and practice: a computer simulation evaluation of best methods," *J. ECT* **22**(3), 169–175 (2006).
32. M. Essenpreis et al., "Spectral dependence of temporal point spread functions in human tissues," *Appl. Opt.* **32**(4), 418–425 (1993).
33. J. C. Ye et al., "NIRS-SPM: statistical parametric mapping for near-infrared spectroscopy," *Neuroimage* **44**(2), 428–447 (2009).
34. M. Brett, I. S. Johnsrude, and A. M. Owen, "The problem of functional localization in the human brain," *Nat. Rev. Neurosci.* **3**, 243–249 (2002).
35. V. S. Fonov et al., "Unbiased nonlinear average age-appropriate brain templates from birth to adulthood," *Neuroimage* **47**, S102–S102 (2009).
36. V. S. Fonov et al., "Unbiased average age-appropriate atlases for pediatric studies," *Neuroimage* **54**(1), 313–327 (2011).
37. D. L. Collins et al., "ANIMAL + INSECT: improved cortical structure segmentation," in *IPMI Lecture Notes Comp. Sci.*, Vol. 1613, pp. 210–223 (1999).
38. X. Cui et al., "A quantitative comparison of NIRS and fMRI across multiple cognitive tasks," *Neuroimage* **54**(4), 2808–2821 (2011).
39. B. Khan et al., "Identification of abnormal motor cortex activation patterns in children with cerebral palsy by functional near infrared spectroscopy," *J. Biomed. Opt.* **15**(3), 036008 (2010).
40. F. Tian et al., "Quantification of functional near infrared spectroscopy to assess cortical reorganization in children with cerebral palsy," *Opt. Express* **18**(25), 25973–25986 (2010).
41. T. Johnstone et al., "Stability of amygdala BOLD response to fearful faces over multiple scan sessions," *Neuroimage* **25**(4), 1112–1123 (2005).
42. S. A. Rombouts et al., "Test-retest analysis with functional MR of the activated area in the human visual cortex," *Am. J. Neuroradiol.* **18**, 1317–1322 (1997).
43. R. A. Charter, "A breakdown of reliability coefficients by test type and reliability method, and the clinical implications of low reliability," *J. Gen. Psychol.* **130**(3), 290–304 (2003).
44. B. Khan et al., "Improving optical contact for functional near-infrared brain spectroscopy and imaging with brush optodes," *Biomed. Opt. Express* **3**(5), 878–898 (2012).
45. L. Becerra et al., "Diffuse optical tomography activation in the somatosensory cortex: Specific activation by painful vs. non-painful thermal stimuli," *PLoS ONE* **4**(11), e8016 (2009).
46. F. A. Kozel et al., "Fractional anisotropy changes after several weeks of daily left high-frequency repetitive transcranial magnetic stimulation of the prefrontal cortex to treat major depression," *J. ECT* **27**(1), 5–10 (2011).
47. R. Chen et al., "Depression of motor cortex excitability by low-frequency transcranial magnetic stimulation," *Neurology* **48**(5), 1398–1403 (1997).
48. H. Nakamura et al., "Intracortical facilitation and inhibition after transcranial magnetic stimulation in conscious humans," *J. Physiol.* **498**(Pt 3), 817–823 (1997).